

CORRIGENDUM-2

Date: 18.03.2020

Ref.: Bid Document No. CUGL/C&P/TEN1920/19,209,055 dated 02.03.2020 for the Procurement of Laser Methane Mini Detectors for Kanpur & Bareilly.

Sub.: Corrigendum-2

Please note the following corrigendum-2

Sr. No.	Clause No.	For	Read As
1	Page No. 87, Clause No. 12	Certificate : ATEX II 2G Ex ib op-pr/op-is IIA T1 Gb, CE or equivalent or Approved as per Indian standard	Certificate : On the basis of ATEX II 2G Ex ib IIA T1 Gb.
2	Page No. 87, Clause No. 14	Units : ppm x m / % vol x m / %LEL x m	Units : ONLY PPM x M required

Note : All other terms & conditions of tender shall remain same.

This corrigendum-2 is to be treated as part of the original bid document and while submitting your offer, corrigendum-2 shall also be signed and stamped along with bid document.

Thanking you

For, Central U.P. Gas Limited


(K.K. Gupta)
Sr. Manager (C&P)